

**Responses to Vendor Questions / Addendum to RFQs 1.4.19**

<b>Question</b>	<b>Answers</b>
<b>General</b>	
--What entity should the quote be made in the name of? Graham County or possibly Smokey Mountain Urgent Care?	All quotes should be made out to Graham County.
--Can we submit a "bundled" bid quote on both pieces of equipment? Or just individual quotes for each piece of equipment.	We will continue to accept individual quotes for each piece of equipment. However, all vendors are welcome to submit a bundled offering on both pieces of equipment.
<b>CT Scanner</b>	
--This question pertains to the inclusion of a Laser Printer. Are you sure you want to include a laser printer? It is an expensive. Most centers utilize PACS (Picture Archive Communications System) for image viewing. Laser printers use film which is expensive and cumbersome to manage.	No, we will not need a laser printer anymore.
--Do you have a PACS at other locations that you could link in to at the new site?	At this time, we do not have any PACS at other locations to link into.
--Lead glass is usually provided by the contractor who does your build out. Please confirm that we need to include that in our proposal.	We will not require lead glass, but a shielding diagram will be required.
--Please confirm that you are requesting a single head injector. Dual head has become an industry standard, and has definite advantage. Especially in vascular related exams. Our original quote contains a dual head system which we can change to a single if required.	We will require all quotes to have a dual head injector.